

**CUSTOMER**

Name: AV-Consulting  
Address: Benedenberg 100A  
Zip code & City : 2861 LH Bergambacht  
Country: The Netherlands

**CALIBRATION OF**

Device: Accelerometer  
Brand & type: Bruel & Kjaer 8305  
Serial number: 1000577  
Customers Instrument tag: AV-039

**SPECIFICATIONS**

Calibrated in accordance: ISO-16063-21  
Method used: Back to Back Calibration, direct comparison to reference standard  
accelerometer as specified in ISO-16063-21  
Traceability: DANAK CDK1602396

**CALIBRATION  
CONDITIONS**

Preconditioning: 4 hours at 23 [°C] ± 3 [°C]  
Environmental conditions: 

Pressure	Unit	Humidity	Unit	Temperature	Unit
1022,1	[hPa]	60,0	[%]	21,7	[°C]

**UNCERTAINTY  
OF  
MEASUREMENT**

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , which for a normal distribution provides a level of confidence of approximately 95%. The uncertainty in evaluation has been carried out in accordance with EA-4/02 from elements originating from standards, calibration methods, effect of environmental conditions and any short time contribution from the device under calibration.

**RESULT** **PASS**

**DATE** Date of calibration : Date of issue :  
Calibration Engineer: Approved Signatory:  
A.Vreeswijk 27-9-2017 27-9-2017

**VISUAL INSPECTION**

Yes No

The equipment / device is in serviceable condition.	x	
There is no visible damage.	x	
The appropriate documentation accompanied the equipment.	x	
Calibration tags / CE tags are present and correct.	x	
The equipment is suitable to use for official testing and/or calibration.	x	

**COMMENTS**

The device under test submitted for testing has successfully completed the periodic tests for the environmental conditions under which the tests were performed.  
 Pass for frequency range and response as specified by manufacturer.

**CALIBRATION EQUIPMENT**

Device	Type	Brand	Serialno.
Digital Voltmeter 6½ digits	34465A	Keysight	MY54502281
Ultra low distortion generator	DS360	Stanford RS	33264
Conditioning Amplifier	Nexus 2692	Bruel & Kjaer	2079137
Laboratory Standard Accelerometer	8305	Bruel & Kjaer	396989
Laboratory Standard Accelerometer	8076K	Kistler	C127545
Power Amplifier	2712	Bruel & Kjaer	2005145
Vibration Exiter	4808	Bruel & Kjaer	2014225

**EXPANDED UNCERTAINTY OF MEASUREMENTS**

Frequency f [Hz]	Expanded Uncertainty, U, Rounded to 2 Significant Digits. Charge Transmission Reference.
3 - 8 Hz	1,3%
10-2000 Hz	1,1%
2500 - 4000 Hz	1,9%
5000 - 8000 Hz	2,6%
10000 Hz	3,4%

Thermal effects have been incorporated in the uncertainties


**MEASUREMENTS**

	Brand	Type	Serial no.	Sensitivity	Unit	Ref. frequenc	Fixture
<b>Z-AXIS</b>	Bruel & Kjaer	8305	1000577	0,1256	pC ms <sup>-2</sup>	159,16	Stud + Oil

**MEASUREMENTS CONDITIONS**

Gravitational Field NL	9,80665	ms <sup>-2</sup>
Excitation Angle	0	Degree
Excitation	Sinusoidal	
Connexion Cable	AO-0038	Customer cable

**Frequentierespons**



**Sensitivity measured relative to reference axis measurement at 159,16 Hz**

f [Hz]	Dev [%]	Dev [dB]
5	0,28	0,04
6,3	0,17	0,03
8	0,17	0,03
10	0,20	0,15
12,5	0,18	0,10
16	0,11	0,04
20	0,23	0,07
25	0,19	0,05
31,5	0,22	0,05
40	0,20	0,04
50	0,11	0,01
63	0,08	0,01
80	0,07	0,01
100	0,01	0,00
125	0,01	0,00
159,16	0,00	0,00
200	0,10	0,01
250	0,09	0,01
315	0,02	0,00

f [Hz]	Dev [%]	Dev [dB]
400	0,00	0,00
500	0,02	0,00
630	0,06	0,01
800	0,09	0,01
1000	0,09	0,01
1250	0,28	0,04
1600	0,26	0,04
2000	0,37	0,05
2500	0,58	0,08
3150	0,98	0,11
4000	1,48	0,14
5000	2,46	0,43
6300	4,19	2,05
8000	6,64	3,83
10000	9,81	3,09
12500	11,83	4,03
16000	14,43	3,84
20000	19,10	4,84